



# Australian Register of Therapeutic Goods Certificate

Issued to

**Stonestar Wholesale Pty Ltd**

for approval to supply

## Severe acute respiratory syndrome-associated coronavirus IVDs

<b>ARTG Identifier</b>	348890
<b>ARTG Start Date</b>	18/11/2020
<b>Product Category</b>	Medical Device Included - IVD Class 3
<b>GMDN</b>	CT772
<b>GMDN Term</b>	Severe acute respiratory syndrome-associated coronavirus IVDs
<b>Intended Purpose</b>	For the rapid, qualitative detection of the nucleocapsid protein antigen from SARS-CoV-2 in human nasal or throat swab specimens.

Manufacturer Details	Address	Certificate number(s)
VivaChek Biotech (Hangzhou) Co Ltd	Level 2 Block 2 146 East Chaofeng Road Yuhang Economy Development Zone Hangzhou , Zhejiang , 311100 China	DV-2020-MC-32312-1

### ARTG Standard Conditions

The above Medical Device Included - IVD Class 3 has been entered on the Register subject to the following conditions:

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

### Products Covered by This Entry

#### 1. Severe acute respiratory syndrome-associated coronavirus IVDs

**This entry:** does not contain System(s)/Procedure Pack(s)

#### IVD Information

Name	Category Description
VivaDiag SARS-CoV-2 Ag Rapid Test	Point of care testing

### Product Specific Conditions

- 1. The person (the sponsor) in relation to whom the Device is included in the ARTG may only supply the Device to a. laboratories that are accredited pathology laboratories and/or b. medical practitioners who are registered under a law of a State or Territory and/or c. health care professionals in residential and aged care facilities and/or d. Commonwealth, State or Territory department of health and/or e. an agency of the Commonwealth, State or Territory acting on behalf of Commonwealth, State or Territory department of health. And within 12 months of an approval the following information will be required to be provided to the TGA. 2. A report of any adverse events, corrective and preventative actions, and customer complaints provided in the context of the number of devices supplied since the introduction of the Device(s) to market in Australia and Worldwide. 3. Information regarding any refusals by Regulatory Authorities for the supply of the Device(s) in any other regulatory jurisdictions. 4. Further analytical and clinical evidence to support a. Analytical and clinical performance of the device b. Device stability (e.g,

shelf-life stability, transport stability) 5. Instructions for use that provide updated information on the analytical and clinical performance characteristics of the device. 6. Evidence of how the user may verify, at the time of use that the device will perform as intended by the manufacturer through the use of controls.

---

Therapeutic Goods Administration  
PO Box 100, Woden ACT 2606 Australia  
Phone: 1800 020 653  
Email: [info@tga.gov.au](mailto:info@tga.gov.au)

ARTG Identifier: 348890  
ARTG Start Date: 18/11/2020